

## TECHNICAL BULLETIN FACTORS AFFECTING FLOW RATE -ON-Q<sup>\*</sup> PAIN RELIEF SYSTEM WITH FIXED FLOW RATE

## HOW THE PUMP WORKS


The ON-Q<sup>\*</sup> pump consists of a dual layer elastomeric membrane with a soft outer protective cover. The pressure that pumps the fluid is generated by the strain energy of the elastomeric membranes that are forced to expand when the pump is filled.

Flow control is achieved with a flow restrictor comprised of a fixed diameter and length in conjunction with the pressurized reservoir. The flow restrictor is a precision bored glass orifice and is located at the distal end of the pre-attached pump tubing.

Elastomeric pumps are calibrated to specific operating conditions. Flow rate accuracy for the  $ON-Q^*$  pump with fixed flow rate is determined under the following conditions: When filled to the labeled volume,  $ON-Q^*$  flow accuracy is  $\pm 15\%$  of the labeled infusion rates when infusion is started within 0-8 hours after fill and delivering normal saline as the diluent at 88° F (31° C) against a back pressure of 16 inches (40 cm).

## FACTORS AFFECTING FLOW RATE

A variety of factors such as fill volume, temperature, pump position, and storage times affect the flow rate accuracy of elastomeric pumps. These factors may result in an increase or decrease in flow rate from the labeled flow rate and impact the delivery time. The information presented on the attached table outlines the factors affecting flow rate and presents information to help ensure accurate delivery times.



MK-00539 rev 4 EN Page 1 of 3 06/2015

## ON-Q\* PAIN RELIEF SYSTEM WITH FIXED FLOW RATE FACTORS AFFECTING FLOW RATE

When filled to the labeled volume,  $ON-Q^*$  flow accuracy is  $\pm 15\%$  of the labeled infusion rates when infusion is started within 0-8 hours after fill and delivering normal saline as the diluent at 88° F (31° C) against a back pressure of 16 inches (40 cm).

	FLOW RATE FACTOR	GUIDELINES
Fill Volume		
	Flow rate accuracy specification is based on filling the pump to the labeled fill volume	Refer to the Delivery Times Table in the product Instructions for Use
	<ul> <li>Filling the pump <u>less</u> than the labeled fill volume ↑ <u>increases</u> flow rate</li> </ul>	• Do not fill less than the minimum fill volume listed on the Delivery Times Table
	• Filling the pump more than the labeled fill volume $\psi$ decreases flow rate	Do not exceed the maximum fill volume
Temperature		
	The flow controller should be in direct contact with the skin (88° F, 31° C). Temperature will affect fluid viscosity:	<ul><li>Tape flow controller to the patient's skin</li><li>Do not place ice, cold therapy or heat in close</li></ul>
	<ul> <li>Flow rate will ↑ <u>increase</u> approximately 1.4% per 1° F/0.6° C increase in</li> </ul>	<ul> <li>proximity to the flow controller</li> <li>If ON-Q* is refrigerated, allow the pump to reach room temperature prior to use</li> </ul>
	<ul> <li>temperature</li> <li>Flow rate will ↓ <u>decrease</u> approximately 1.4% per 1° F/0.6° C decrease in temperature</li> </ul>	<ul> <li>Instruct patient not to place pump underneath bed covers where it may become too warm</li> </ul>
Pump Position		
ON-O Dir Reite Verte Watan	The pump should be positioned approximately 16 inches (40 cm) below the level of the catheter site	Do not place the pump on the floor or hang from a bed post or IV pole
	<ul> <li>Positioning the pump above this level may ↑ increase the flow rate</li> <li>Positioning the pump below this level may ↓ decrease flow rate</li> </ul>	• Use E-clip to clip to the patient's dressing or clothing or place the pump in the carrying case
Storage Time		
	<ul> <li>Flow rate accuracy specification is based on starting the infusion within 0-8 hours after filling</li> <li>Pumps stored for more than 8 hours prior to starting infusion may ↓ decrease flow rate</li> </ul>	• When a filled pump is stored beyond 8 hours before use, the pressure in the reservoir will decrease due to stretch of the elastomeric membranes, which may result in a reduction in flow rate below the labeled rate
		Technical Bulletin (Effect of Storage Times on Flow Rate of Pre-Filled ON-Q* Elastomeric Pumps) available at www.avanospainmanagement.com
External Pressure		
-	<ul> <li>Squeezing or laying on the pump may         <u>↑ increase</u> flow rate</li> </ul>	• Do not squeeze the pump. The pump has sufficient force to infuse the medication
ON-Q		<ul> <li>Instruct patient to place the pump on a bedside table or other location when sleeping, to help prevent laying on the pump for extended periods of time</li> </ul>

Note: If the ON-Q\* device did not perform as expected, do not discard. Contact Avanos Medical for product return instructions.

There are inherent risks in all medical devices. Please refer to the product labeling for **Indications, Cautions, Warnings and Contraindications**. Failure to follow the product labeling could directly impact patient safety. Physician is responsible for prescribing and administering medications per instructions provided by the drug manufacturer. Refer to **www.avanospainmanagement.com** for additional product safety **Technical Bulletins**.

Please contact the Clinical Services Department at **800-444-2728** or **949-923-2400** if you have any questions regarding this information.



Rx only. \*Registered Trademark or Trademark of Avanos Medical, Inc., or its affiliates. © 2018 AVNS. All rights reserved. MK-00539 rev4 EN 06/15